

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Scope of accreditation

#### Calibration according to SS-EN ISO/IEC 17025:2018

RISE Research Institutes of Sweden AB

Borås

Accreditation number

1002

Kontroll och kalibrering

A002626-012

### Electricity and Magnetism

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Capacitance	Inhouse method; KVf 20		Capacitance showing	0,22 to 3,2999 nF	0,39 % + 0,0078 nF		Yes	2	Yes	
			Capacitance showing	0,33 to 0,46999 mF	0,5 % + 160 nF		Yes	2	Yes	
			Capacitance showing	0,33 to 0,579999 µF	0,3 % + 0,16 nF		Yes	2	Yes	
			Capacitance showing	0,47 to 1,09999 mF	0,35 % + 780 nF		Yes	2	Yes	
			Capacitance showing	0,58 to 1,09999 µF	0,19 % + 0,78 nF		Yes	2	Yes	
			Capacitance showing	1,1 to 2,8 mF	0,39 % + 1,2 µF		Yes	2	Yes	
			Capacitance showing	1,1 to 3,29999 µF	0,31 % + 1,2 nF		Yes	2	Yes	
			Capacitance showing	11 to 32,9999 µF	0,31 % + 23 nF		Yes	2	Yes	
			Capacitance showing	11 to 32,9999 mF	0,58 % + 23 µF		Yes	2	Yes	
			Capacitance showing	110 to 280 µF	0,39 % + 120 nF		Yes	2	Yes	
			Capacitance showing	110 to 329,999 nF	0,19 % + 0,023 nF		Yes	2	Yes	
			Capacitance showing	2,800001 to 3,29999 mF	0,35 % + 2,3 µF		Yes	2	Yes	
			Capacitance showing	280,001 to 329,999 µF	0,35 % + 230 nF		Yes	2	Yes	
			Capacitance showing	3,3 to 10,9999 µF	0,19 % + 7,8 nF		Yes	2	Yes	
			Capacitance showing	3,3 to 109,999 nF	0,19 % + 0,0078 nF		Yes	2	Yes	
			Capacitance showing	3,3 to 4,6999 mF	0,5 % + 1,6 µF		Yes	2	Yes	
			Capacitance showing	33 to 110 mF	0,78 % + 47 µF		Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Electricity and Magnetism

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Capacitance	Inhouse method; KVf 20		Capacitance showing	33 to 40,699 µF	0,5 % + 16 nF		Yes	2	Yes	
			Capacitance showing	4,7 to 10,9999 mF	0,35 % + 7,8 µF		Yes	2	Yes	
			Capacitance showing	40,7 to 109,999 µF	0,35 % + 78 nF		Yes	2	Yes	
Conductivity	Inhouse method; KVf 20		conductance showing	0,9 to 2,99999 µS	0,0026 %		Yes	2	Yes	
			conductance showing	2,5 to 2,99999 nS	0,4 %		Yes	2	Yes	
			conductance showing	3 to 8,99999 nS	0,26 %		Yes	2	Yes	
			conductance showing	30 to 89,999 nS	0,025 %		Yes	2	Yes	
			conductance showing	300 to 899,99 nS	0,0054 %		Yes	2	Yes	
			conductance showing	9 to 29,9999 nS	0,041 %		Yes	2	Yes	
			conductance showing	90 to 299,99 nS	0,01 %		Yes	2	Yes	
Current	Inhouse method; KVf 20	AC	Current showing	0,33 to 1,09999 A	0,039 % ppm + 78 µA		Yes	2	Yes	45 Hz – 1 kHz
		AC	Current showing	0,33 to 2,49999 mA	0,078 % ppm + 0,12 µA		Yes	2	Yes	45 Hz – 1 kHz
		AC	Current showing	1,1 to 2,99999 A	0,047 % ppm + 78 µA		Yes	2	Yes	45 Hz – 1 kHz
		AC	Current showing	11 to 20 A	0,093 % ppm + 3900 µA		Yes	2	Yes	45 Hz – 100 Hz
		AC	Current showing	2,5 to 3,29999 mA	0,07 % ppm + 0,3 µA		Yes	2	Yes	45 Hz – 1 kHz
		AC	Current showing	20,001 to 29,999 A	0,3 % + 1,6 mA		Yes	2	Yes	45 Hz – 440 Hz, current coil
		AC	Current showing	200,001 to 1000 A	0,3 % + 156 mA		Yes	2	Yes	45 Hz – 100 Hz, current coil
		AC	Current showing	29 to 329,99 µA	0,097 % ppm + 0,078 µA		Yes	2	Yes	45 Hz – 1 kHz

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Electricity and Magnetism

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Current	Inhouse method; KVf 20	AC	Current showing	3 to 10,9999 A	0,039 % ppm + 1550 µA		Yes	2	Yes	45 Hz – 100 Hz
		AC	Current showing	3,3 to 32,9999 mA	0,031 % ppm + 1,6 µA		Yes	2	Yes	45 Hz – 1 kHz
		AC	Current showing	30 to 32 A	0,3 % + 32 mA		Yes	2	Yes	45 Hz – 100 Hz, current coil
		AC	Current showing	32,001 to 200 A	0,3 % + 32 mA		Yes	2	Yes	45 Hz – 100 Hz, current coil
		AC	Current showing	33 to 329,999 mA	0,031 % ppm + 16 µA		Yes	2	Yes	45 Hz – 1 kHz
		DC	Current showing	+/-1,09999 A	155 ppm + 31 µA		Yes	2	Yes	
		DC	Current showing	+/-10,9999 A	390 ppm + 390 µA		Yes	2	Yes	
		DC	Current showing	+/-1000 A	0,26 % + 225 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-105 A	0,24 % + 3,9 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-109,999 A	0,24 + 3,9 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-17,3999 A	775 ppm + 580 µA		Yes	2	Yes	
		DC	Current showing	+/-189,999 µA	140 ppm + 0,011 µA		Yes	2	Yes	
		DC	Current showing	+/-2,99999 A	295 ppm + 31 µA		Yes	2	Yes	
		DC	Current showing	+/-20 A	550 ppm + 4500 µA		Yes	2	Yes	
		DC	Current showing	+/-200 A	0,26 % + 45 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-3,29999 mA	78 ppm + 0,038 µA		Yes	2	Yes	
		DC	Current showing	+/-32 A	0,24 % + 3,9 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-32,9999 mA	78 ppm + 0,19 µA		Yes	2	Yes	
		DC	Current showing	+/-329,999 µA	120 ppm + 0,016 µA		Yes	2	Yes	

**Appendix 1**

Date

2023-07-05

Reference

2022/873

**Electricity and Magnetism**

<b>Technology area</b>	<b>Method</b>	<b>Parameter</b>	<b>Material</b>	<b>Measure</b>	<b>Best measuring ability (CMC) +/-</b>	<b>Technique</b>	<b>Flex</b>	<b>Type of flex</b>	<b>Field</b>	<b>Note</b>
Current	Inhouse method; KVf 20	DC	Current showing	+/-329,999 mA	78 ppm + 1,9 µA		Yes	2	Yes	
		DC	Current showing	+/-525 A	0,24 % + 20 mA		Yes	2	Yes	Current coil
		DC	Current showing	+/-549,999 A	0,24 % + 20 mA		Yes	2	Yes	Current coil
Resistance	Inhouse method; KVf 20	ESR	Resistance showing	0 ohm	0,00078 Ohm		Yes	2	Yes	
		ESR	Resistance showing	0,0011 to 10,999 Ohm	31 ppm + 0,00078 Ohm		Yes	2	Yes	
		ESR	Resistance showing	0,110 to 1,099999 kOhm	22 ppm + 0,0016 Ohm		Yes	2	Yes	
		ESR	Resistance showing	0,110 to 1,099999 MOhm	25 ppm + 1,6 Ohm		Yes	2	Yes	
		ESR	Resistance showing	1,1 to 10,99999 kOhm	22 ppm + 0,016 Ohm		Yes	2	Yes	
		ESR	Resistance showing	1.1 to 3,299999 MOhm	47 ppm + 23 Ohm		Yes	2	Yes	
		ESR	Resistance showing	11 to 109,9999 kOhm	22 ppm + 0,16 Ohm		Yes	2	Yes	
		ESR	Resistance showing	11 to 32,9999 Ohm	23 ppm + 0,0012 Ohm		Yes	2	Yes	
		ESR	Resistance showing	11 to 32,99999 MOhm	190 ppm + 1900 Ohm		Yes	2	Yes	
		ESR	Resistance showing	110 to 137 MOhm	2600 ppm + 40000 Ohm		Yes	2	Yes	
		ESR	Resistance showing	137,001 to 400 MOhm	2300 ppm + 78000 Ohm		Yes	2	Yes	
		ESR	Resistance showing	3,3 to 10,99999 MOhm	100 ppm + 39 Ohm		Yes	2	Yes	
		ESR	Resistance showing	33 to 109,9999 MOhm	390 ppm + 2300 Ohm		Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Electricity and Magnetism

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Resistance	Inhouse method; KVf 20	ESR	Resistance showing	33 to 109,9999 Ohm	22 ppm + 0,0011 Ohm		Yes	2	Yes	
		ESR	Resistance showing	400,001 to 1100 MOhm	11600 ppm + 39000 Ohm					
Voltage	Inhouse method; KVf 20	AC	Voltage showing	0,33 to 3,29999 V	120 ppm + 47 µV		Yes	2	Yes	45 Hz – 10 kHz
		AC	Voltage showing	1,000 to 32,999 mV	120 ppm + 4,7 µV					
		AC	Voltage showing	3,3 to 32,9999 V	120 ppm + 470 µV		Yes	2	Yes	45 Hz – 10 kHz
		AC	Voltage showing	33 to 329,999 mV	110 ppm + 6,2 µV					
		AC	Voltage showing	33 to 329,999 V	150 ppm + 1550 µV		Yes	2	Yes	45 Hz – 1 kHz
		AC	Voltage showing	330 to 1050 V	190 ppm + 7800 µV					
		DC	Voltage showing	+/- 1050 V	14 ppm+ 1160 µV		Yes	2	Yes	
		DC	Voltage showing	+/- 3,299999 V	9 ppm + 1,6 µV					
		DC	Voltage showing	+/- 32,99999 V	9 ppm + 16 µV		Yes	2	Yes	
		DC	Voltage showing	+/- 329,9999 mV	16 ppm + 0,78 µV					
		DC	Voltage showing	+/- 329,9999 V	14 ppm + 116 µV		Yes	2	Yes	

### Length related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Length	Inhouse method; KVf 30		dial guage	0 - 1	7 µm		Yes	2	Yes	
			Measuring gauge	0 - 30 mm	7 µm					
	Inhouse method; KVf 31		Micrometer	0 - 1000 mm	3 µm		Yes	2	Yes	
			Three point micrometer	6 - 150 mm	4 µm					
	Inhouse method; KVf 32		Calliper	0 - 1500 mm	20 µm		Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Length related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Length	Inhouse method; KVj 62		Extensometer	0 - 60 mm	0,30 µm		Yes	2	Yes	

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Flow	Inhouse method; 2527	Gas flow	Gas flow meter	0,00001 – 120 g/s	0,9% of actual flow		Yes	2	Yes	
		Gas flow	Gas flow meter	0,12 – 2,9 kg/s	1,3-1,4% of actual flow		Yes	2	Yes	
	NT VVS 018	Air flow/Air speed	Anemometer	0,1 – 0,5 m/s	0,02 m/s		Yes	2	No	
		Air flow/Air speed	Anemometer	0,5 – 1,0 m/s	0,03 m/s		Yes	2	No	
		Air flow/Air speed	Anemometer	1,0 – 4,0 m/s	2,9% of actual speed (m/s)		Yes	2	No	
		Air flow/Air speed	Anemometer	4,0 – 35,0 m/s	3,7% of actual speed (m/s)		Yes	2	No	
Force	Inhouse method; KVj 60		Tensile testing machine	1N - 1MN	0,12 %		Yes	2	Yes	
			Tension and pressure testing machines	1N - 5MN	0,12 %		Yes	2	Yes	
	SS-EN ISO 7500-1		Tensile testing machine	1N - 1MN	0,12 %		Yes	2	Yes	
			Tension and pressure testing machines	1N - 5MN	0,12 %		Yes	2	Yes	
Impact testing	Inhouse method; KVj 61		Impact testing machine	1J - 200J	1 J	Indirekt	Yes	2	Yes	
Mass	Inhouse method; KVj 18		NAWI	0,1-1 g	2-3 µg	E1	Yes	2	Yes	
			NAWI	0,5-1 kg	0,26-0,76 mg	E2	Yes	2	Yes	
			NAWI	100-200 g	17-30 µg	E1	Yes	2	Yes	
			NAWI	100-200 kg	0,58-1,2 g	F1	Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Mass	Inhouse method; KVj 18		NAWI	10-100 g	6-17 µg	E1	Yes	2	Yes	
			NAWI	10-100 kg	0,017-0,58 g	F1	Yes	2	Yes	
			NAWI	10-100 mg	1-2 µg	E1	Yes	2	Yes	
			NAWI	1-10 g	3-6 µg	E1	Yes	2	Yes	
			NAWI	1-10 kg	0,76-6,2 mg	E2	Yes	2	Yes	
			NAWI	1-10 mg	1-1 µg	E1	Yes	2	Yes	
			NAWI	200-500 g	30-73 µg	E1	Yes	2	Yes	
	Inhouse method; KVj 19		Automatic weighing instruments	10000-20000 kg	2-4 kg	M2	Yes	2	Yes	
			Automatic weighing instruments	1000-5000 kg	0,2-1 kg	M2	Yes	2	Yes	
			Automatic weighing instruments	100-500 kg	9-41 g	M1	Yes	2	Yes	
			Automatic weighing instruments	1-100 kg	6-9 g	M1	Yes	2	Yes	
			Automatic weighing instruments	5000-10000 kg	1-2 kg	M2	Yes	2	Yes	
			Automatic weighing instruments	500-1000 kg	0,1-0,2 kg	M2	Yes	2	Yes	
	Inhouse method; KVj 41		Weight	100 kg	3 g	M1 – M2	Yes	2	Yes	
			Weight	1000 kg	30 g	M1 – M2	Yes	2	Yes	
			Weight	500 kg	10 g	M1 – M2	Yes	2	Yes	
	Inhouse method; KVj 44		NAWI	0,5-1000 kg	0,3-82 g	M1	Yes	2	Yes	
			NAWI	100-300 ton	35-81 kg	M2	Yes	2	Yes	
			NAWI	10-100 ton	3,5-35 kg	M2	Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Mass	Inhouse method; KVj 44		NAWI	1-10 ton	0,35-3,5 kg	M2	Yes	2	Yes	
	Inhouse method; KVj 6	Weight		1 g	0,3 mg	M1	Yes	2	Yes	
		Weight		1 kg	16 mg	M1	Yes	2	Yes	
		Weight		1 mg	0,06 mg	M1	Yes	2	Yes	
		Weight		10 g	0,6 mg	M1	Yes	2	Yes	
		Weight		10 kg	160 mg	M1	Yes	2	Yes	
		Weight		10 mg	0,08 mg	M1	Yes	2	Yes	
		Weight		100 g	1,6 mg	M1	Yes	2	Yes	
		Weight		100 mg	0,16 mg	M1	Yes	2	Yes	
		Weight		2 g	0,4 mg	M1	Yes	2	Yes	
		Weight		2 kg	30 mg	M1	Yes	2	Yes	
		Weight		2 mg	0,06 mg	M1	Yes	2	Yes	
		Weight		20 g	0,8 mg	M1	Yes	2	Yes	
		Weight		20 kg	300 mg	M1	Yes	2	Yes	
		Weight		20 mg	0,10 mg	M1	Yes	2	Yes	
		Weight		200 g	3,0 mg	M1	Yes	2	Yes	
		Weight		200 mg	0,20 mg	M1	Yes	2	Yes	
		Weight		5 g	0,5 mg	M1	Yes	2	Yes	
		Weight		5 kg	80 mg	M1	Yes	2	Yes	
		Weight		5 mg	0,06 mg	M1	Yes	2	Yes	
		Weight		50 g	1,0 mg	M1	Yes	2	Yes	
		Weight		50 kg	800 mg	M1	Yes	2	Yes	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Mass	Inhouse method; KVj 6		Weight	50 mg	0,12 mg	M1	Yes	2	Yes	
			Weight	500 g	8,0 mg	M1	Yes	2	Yes	
			Weight	500 mg	0,25 mg	M1	Yes	2	Yes	
	Inhouse method; KVj 7		Weight	1 g	0,10 mg	M1	Yes	2	No	
			Weight	1 kg	5,0 mg	M1	Yes	2	No	
			Weight	1 mg	0,02 mg	M1	Yes	2	No	
			Weight	10 g	0,20 mg	M1	Yes	2	No	
			Weight	10 kg	50 mg	M1	Yes	2	No	
			Weight	10 mg	0,03 mg	M1	Yes	2	No	
			Weight	100 g	0,5 mg	M1	Yes	2	No	
			Weight	100 mg	0,05 mg	M1	Yes	2	No	
			Weight	2 g	0,12 mg	M1	Yes	2	No	
			Weight	2 kg	10 mg	M1	Yes	2	No	
			Weight	2 mg	0,02 mg	M1	Yes	2	No	
			Weight	20 g	0,25 mg	M1	Yes	2	No	
			Weight	20 kg	100 mg	M1	Yes	2	No	
			Weight	20 mg	0,03 mg	M1	Yes	2	No	
			Weight	200 g	1,0 mg	M1	Yes	2	No	
			Weight	200 mg	0,06 mg	M1	Yes	2	No	
			Weight	5 g	0,16 mg	M1	Yes	2	No	
			Weight	5 kg	25 mg	M1	Yes	2	No	
			Weight	5 mg	0,02 mg	M1	Yes	2	No	

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Mass	Inhouse method; KVj 7		Weight	50 g	0,3 mg	M1	Yes	2	No	
			Weight	50 kg	250 mg	M1	Yes	2	No	
			Weight	50 mg	0,04 mg	M1	Yes	2	No	
			Weight	500 g	2,5 mg	M1	Yes	2	No	
			Weight	500 mg	0,08 mg	M1	Yes	2	No	
	Inhouse method; KVj 70		Vehicle gas dispenser	< 18,5 kg/min	0,55 %	Mass flow meter	Yes	2	Yes	
			Automatic weighing instruments	100-500 kg	6-60 g	M1	Yes	2	Yes	
			Automatic weighing instruments	10-100 kg	0,6-6 g	F1	Yes	2	Yes	
			Automatic weighing instruments	1-10 kg	6-60 mg	E2	Yes	2	Yes	
			Automatic weighing instruments	1-1000 g	0,6-6 mg	E2	Yes	2	Yes	
Pressure	Inhouse method; 3635	Gauge pressure	Press showing measuring instrument	±20 kPa	0.54% of actual pressure, but not less than 0.7 Pa		Yes	2	Yes	
		Gauge pressure	Press showing measuring instrument	±23000 Pa	0.2% of actual pressure, but not less than 0.3 Pa		Yes	2	No	
	Inhouse method; KVf 34	Absolute pressure	Press showing measuring instrument	900-1100 mbar	1,1 hPa		Yes	2	Yes	Gas
		Gauge pressure	Press showing measuring instrument	> 2 MPa - 7 MPa	1,4 kPa		Yes	2	Yes	Gas
		Gauge pressure	Press showing measuring instrument	> 20 MPa - 70 MPa	50 kPa		Yes	2	Yes	Vatten

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Mass related quantities

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Pressure	Inhouse method; KVf 34	Gauge pressure	Press showing measuring instrument	> 200 kPa - 2 MPa	0,4 kPa		Yes	2	Yes	Gas
		Gauge pressure	Press showing measuring instrument	> 7 MPa - 20 MPa	4 kPa		Yes	2	Yes	Gas
		Gauge pressure	Press showing measuring instrument	-95 kPa - 200 kPa	0,04 kPa		Yes	2	Yes	Gas
Torque	Inhouse method; KVf 35		Torque wrench	> 10 Nm – 60 Nm	0,18 Nm		Yes	2	Yes	
			Torque wrench	> 5 Nm – 10 Nm	0,09 Nm		Yes	2	Yes	
			Torque wrench	> 500 Nm – 1500 Nm	10 Nm		Yes	2	Yes	
			Torque wrench	> 60 Nm – 500 Nm	1,9 Nm		Yes	2	Yes	
			Torque wrench	0,5 Nm - 5 Nm	0,011 Nm		Yes	2	Yes	
Volume	Inhouse method; KVj 73		Volume meters	< 60 l/min	0,11 %		Yes	2	Yes	10L and 20L handheld standard
			Volume meters	< 60 l/min	0,11 %		Yes	2	Yes	10L and 20L Equipment permanently mounted on trailer
			Volume meters	< 60 l/min	0,12 %		Yes	2	Yes	5L handheld
			Volume meters	> 60 l/min	0,13 %		Yes	2	Yes	From 50L to 2000L . Equipment permanently mounted on trailer

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Humidity	Inhouse method; 2664	Dew point	Humidity meter	-17 - +62°C	0,16 °C		Yes	2	No	
		Dew point	Humidity meter	-55 - +2°C	0,12 °C		Yes	2	No	
		Frost point	Humidity meter	-15 - +0°C	0,16 °C		Yes	2	No	
		Frost point	Humidity meter	-50 - +0°C	0,12 °C		Yes	2	No	
		Relative humidity	Humidity meter	1 – 95 %-rh	1.0 %-rh + 1.3 % of true relative vapor content, but at least 1.3 %-rh		Yes	2	Yes	
		Relative humidity	Humidity meter	1 – 95 %-rh	1.43% of true relative vapor content, but at least 0.25%-rh		Yes	2	No	
Temperature	Inhouse method; 2664		Temperature sensors	-15 - +62°C	0,15 °C		Yes	2	No	
			Temperature sensors	-20 - +80°C	0,3°C + 1,0% of actual temperature difference against ambient		Yes	2	Yes	
	Inhouse method; KVf 25		Contact thermometer	200 °C - 600 °C	2 °C		Yes	2	Yes	Nordtest-method NT VVS 102 & 103
			Contact thermometer	25 °C – 95 °C	0,07 °C		Yes	2	Yes	Nordtest-method NT VVS 102 & 103
			Contact thermometer	-30 °C - 25 °C	0,13 °C		Yes	2	Yes	Nordtest-method NT VVS 102 & 103
			Contact thermometer	95 °C - 200 °C	0,09 °C		Yes	2	Yes	Nordtest-method NT VVS 102 & 103
	Inhouse method; KVf 26		Pyrometer	-15 °C - 120 °C	2,7 °C		Yes	2	Yes	4180, 4181 Precision infrared Calibrator Users Guide.

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type B	1550,1 °C to 1820,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type B	600,0 °C to 800,0°C	0,28 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type B	800,1 °C to 1550,0 °C	0,22 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type C	0,0 °C to 1000,0 °C	0,13 °C		Yes	2	Yes	ASTM standard E 988-96
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type C	1000,1 °C to 1800,0 °C	0,18 °C		Yes	2	Yes	ASTM standard E 988-96
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type C	1800,1 °C to 2000,0 °C	0,21 °C		Yes	2	Yes	ASTM standard E 988-96
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type C	2000,1 °C to 2316,0 °C	0,28 °C		Yes	2	Yes	ASTM standard E 988-96
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type E	0,1 °C to 600,0 °C	0,07 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type E	-199,9 °C to -100,0 °C	0,10 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type E	-250,0 °C to -200,0 °C	0,20 °C		Yes	2	Yes	IEC-584

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type E	600,1 °C to 1000,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type E	-99,9 °C to 0,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type J	-210,0 °C to -100,0 °C	0,11 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type J	800,1 °C to 1200,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type J	-99,9 °C to 800,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type K	-199,9 °C to -100,0 °C	0,13 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type K	-250,0 °C to -200,0 °C	0,36 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type K	500,1 °C to 800,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type K	800,1 °C to 1372,0 °C	0,11 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type K	-99,9 °C to 500,0 °C	0,08 °C		Yes	2	Yes	IEC-584

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type L	-200,0 °C to -100,0 °C	0,08 °C		Yes	2	Yes	DIN 43710-1985
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type L	-99,9 °C to 900,0 °C	0,08 °C		Yes	2	Yes	DIN 43710-1985
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	0,1 °C to 100,0 °C	0,09 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	100,1 °C to 800,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	-199,9 °C to -100,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	-250,0 °C to -200,0 °C	0,57 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	800,1 °C to 1300,0 °C	0,10 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type N	-99,9 °C to 0,0 °C	0,10 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	0,1 °C to 100,0 °C	0,31 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	100,1 °C to 400,0 °C	0,22 °C		Yes	2	Yes	IEC-584

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	1000,1 °C to 1600,0 °C	0,15 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	1600,1 °C to 1767,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	-24,9 °C to 0,0 °C	0,35 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	400,1 °C to 600,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	-50,0 °C to -25,0 °C	0,43 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type R	600,1 °C to 1000,0 °C	0,17 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	0,1 °C to 100,0 °C	0,30 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	100,1 °C to 400,0 °C	0,23 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	1000,1 °C to 1600,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	1600,1 °C to 1767,0 °C	0,21 °C		Yes	2	Yes	IEC-584

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	-24,9 °C to 0,0 °C	0,34 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	400,1 °C to 600,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	-50,0 °C to -25,0 °C	0,40 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type S	600,1 °C to 1000,0 °C	0,18 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type T	0,1 °C to 200,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type T	-199,9 °C to -100,0 °C	0,13 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type T	200,1 °C to 400,0 °C	0,08 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type T	-250,0 °C to -200,0 °C	0,28 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type T	-99,9 °C to 0,0 °C	0,09 °C		Yes	2	Yes	IEC-584
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type U	0,1 °C to 200,0 °C	0,08 °C		Yes	2	Yes	DIN 43710-1985

## Appendix 1

Date

2023-07-05

Reference

2022/873

### Temperature

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Temperature	Inhouse method; KVf 27	Electrically simulated temperature	Temperature simulating instrument, thermocouple type U	-200,0 °C to 0,0 °C	0,13 °C		Yes	2	Yes	DIN 43710-1985
		Electrically simulated temperature	Temperature simulating instrument, thermocouple type U	200,1 °C to 600,0 °C	0,08 °C		Yes	2	Yes	DIN 43710-1985
	Inhouse method; KVf 28		Climate cabinet	-30 °C to 200 °C	0,5 °C		Yes	2	Yes	IEC-584

### Time and Frequency

Technology area	Method	Parameter	Material	Measure	Best measuring ability (CMC) +/-	Technique	Flex	Type of flex	Field	Note
Frequency	Inhouse method; KVf 20		Frequency showing	0,01 to 0,49 Hz	1,9 ppm + 0,0039 Hz		Yes	2	Yes	
			Frequency showing	0,5 Hz to 10 MHz	0,25 ppm		Yes	2	Yes	
		Duty cycles	Pulse showing	1 to 99 %	0,01 %		Yes	2	Yes	

Calibration and measurement capability, CMC, is the smallest uncertainty the calibration laboratory can provide, expressed as the expanded uncertainty having a coverage probability of approximately 95%.

Changes in the scope of accreditation are in bold.

The scope of accreditation is flexible as specified in this decision. The accredited body must always retain a current list of the scope for which it is accredited.

#### Type of flexible scope

1: - Introduce new version of standard method and make editorial changes to non-standard method

2: - Introduce new version of standard method and make editorial changes to non-standard method - Introduce new version and modifications of non-standard method. The procedure must be equivalent