

Appendix 1

Date

2024-02-01

Reference

2023/1665

Scope of accreditation

Calibration according to SS-EN ISO/IEC 17025:2018

Temperature Sensitive Solutions Systems Sweden Solna
 AB

Accreditation number

10592

A014990-001

Temperature

<i>Technology area</i>	<i>Method</i>	<i>Material</i>	<i>Measure</i>	<i>Best measuring ability (CMC) +/-</i>	<i>Technique</i>	<i>Field</i>
Temperature	Inhouse method; CAL SOP 1 version F	Temperature sensor	-80°C- +85°C	0,1 °C		No

Changes in the scope of accreditation are in bold.